
Midmark M11 - v1.0.5
Total Cycles: 2166

Name: RITTER2
Sterilizer ID: V2313228
03/22/2023 10:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.2 F 0.0
2:00 203.7 F 3.5
4:00 217.9 F 9.3
6:00 228.1 F 7.3
8:00 236.1 F 11.4
10:00 236.5 F 9.5
12:00 238.1 F 9.0
14:00 250.5 F 16.0
16:00 263.3 F 23.4
18:00 271.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.5 F 29.6
2:00 271.5 F 28.8
3:00 272.5 F 29.8
4:00 272.0 F 29.2

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 226.7 F 0.0
10:00 237.0 F 0.0
15:00 236.5 F 0.0
20:00 230.8 F 0.0
25:00 228.6 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:16
CYCLE COMPLETE

03/22/2023 11:22 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____