
Midmark M11 - v1.0.5
Total Cycles: 1589

Name: RITTER2
Sterilizer ID: V2313228
10/12/2022 05:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.1 F 0.0
2:00 193.8 F 4.2
4:00 209.6 F 9.3
6:00 226.4 F 9.3
8:00 234.2 F 10.8
10:00 232.6 F 8.5
12:00 238.5 F 10.5
14:00 247.0 F 14.1
16:00 258.7 F 21.0
18:00 269.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.9
1:00 272.6 F 30.1
2:00 272.7 F 30.5
3:00 272.0 F 30.0
4:00 272.6 F 30.2

Min 272.0 F 29.8
Max 272.9 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 235.3 F 0.0
10:00 231.3 F 0.0
15:00 218.8 F 0.0
20:00 217.1 F 0.0
25:00 228.6 F 0.0
30:00 229.0 F 0.0

Drying Complete

FILLING: 2:11
HEATING: 19:50
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

10/12/2022 06:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____