

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2410

Name: RITTER2  
Sterilizer ID: V2313228  
05/24/2023 02:54 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
2:00 177.8 F 3.8  
4:00 191.4 F 8.4  
6:00 212.7 F 9.6  
8:00 227.2 F 9.4  
10:00 232.5 F 8.8  
12:00 243.0 F 13.3  
14:00 243.4 F 12.5  
16:00 256.7 F 19.6  
18:00 268.5 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.6 F 29.8  
2:00 271.6 F 29.0  
3:00 272.1 F 29.9  
4:00 272.4 F 29.6  
  
Min 271.2 F 28.9  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
5:00 232.1 F 0.0  
10:00 223.2 F 0.0  
15:00 222.6 F 0.0  
20:00 222.7 F 0.0  
25:00 222.3 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 19:10  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

05/24/2023 03:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_