

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2904

Name: RITTER2  
Sterilizer ID: V2313228  
09/19/2023 11:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.1 F 0.0  
2:00 168.7 F 1.8  
4:00 182.5 F 5.6  
6:00 208.6 F 10.2  
8:00 221.5 F 6.5  
10:00 233.2 F 9.9  
12:00 232.2 F 8.9  
14:00 242.4 F 12.5  
16:00 241.4 F 11.3  
18:00 254.1 F 17.7  
20:00 265.0 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.8  
1:00 272.6 F 29.4  
2:00 271.8 F 29.2  
3:00 272.1 F 29.2  
4:00 272.6 F 29.5

Min 271.0 F 28.6  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
5:00 197.3 F 0.0  
10:00 198.2 F 0.0  
15:00 194.5 F 0.0  
20:00 190.1 F 0.0  
25:00 186.8 F 0.0  
30:00 183.7 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 21:56  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:01:17  
CYCLE COMPLETE

09/19/2023 12:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_