
Midmark M11 - v1.0.0
Total Cycles: 1738

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2022 08:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 214.5 F 0.0
2:00 195.2 F 1.5
4:00 194.5 F 5.7
6:00 211.1 F 11.3
8:00 227.1 F 6.5
10:00 240.4 F 11.9
12:00 237.4 F 9.4
14:00 238.2 F 9.1
16:00 248.9 F 14.5
18:00 261.1 F 21.4
20:00 272.1 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.6
1:00 272.2 F 28.9
2:00 272.5 F 29.2
3:00 272.7 F 29.2
4:00 272.7 F 29.2

Min 271.7 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 233.5 F 0.0
10:00 225.3 F 0.0
15:00 233.3 F 0.0
20:00 226.8 F 0.0
25:00 230.8 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 21:01
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:60
CYCLE COMPLETE

03/24/2022 09:19 AM
ACCEPTED: X
REJECTED:
Operator: U02