

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4665

Name: RITTER 1  
Sterilizer ID: V2178198  
06/20/2024 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 104.9 F 0.0  
2:00 105.7 F 1.0  
4:00 130.0 F 3.3  
6:00 162.7 F 6.6  
8:00 192.1 F 10.7  
10:00 206.6 F 2.8  
12:00 218.7 F 7.4  
14:00 230.3 F 11.7  
16:00 230.7 F 7.4  
18:00 242.6 F 12.9  
20:00 241.0 F 10.9  
22:00 253.2 F 17.0  
24:00 264.2 F 23.4  
26:00 271.5 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 28.8  
2:00 271.7 F 28.6  
3:00 272.4 F 28.7  
4:00 271.6 F 28.6

Min 271.4 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
5:00 215.2 F 0.0  
10:00 218.9 F 0.0  
15:00 229.3 F 0.0  
20:00 229.2 F 0.0  
25:00 226.1 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 26:00  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:04:17  
CYCLE COMPLETE

06/20/2024 12:49 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_