
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 005207
17:06 09/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:56
STERILIZING 2:47
136.0 C 220kPa 2:47
136.1 C 222kPa 3:17
136.3 C 220kPa 3:47
136.2 C 220kPa 4:17
136.4 C 224kPa 4:47
136.4 C 221kPa 5:17
136.2 C 221kPa 5:47
136.1 C 222kPa 6:17
136.1 C 222kPa 6:17
VENTING 6:17
AIR DRYING 8:10
CYCLE COMPLETE 10:45

Digital Signature #
6480694B92D9E1E2
