

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3365

Name: RITTER 1  
Sterilizer ID: V2178198  
06/09/2023 09:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.8 F 0.0  
2:00 187.9 F 0.2  
4:00 199.4 F 4.9  
6:00 215.0 F 9.4  
8:00 225.7 F 8.7  
10:00 233.0 F 9.5  
12:00 238.0 F 10.1  
14:00 243.7 F 12.5  
16:00 242.0 F 11.2  
18:00 254.4 F 17.5  
20:00 265.1 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.5 F 28.9  
2:00 271.8 F 28.2  
3:00 271.8 F 28.6  
4:00 272.5 F 28.8  
  
Min 271.4 F 28.0  
Max 272.6 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.0 F 0.0  
5:00 204.1 F 0.0  
10:00 207.1 F 0.0  
15:00 205.2 F 0.0  
20:00 206.2 F 0.0  
25:00 208.1 F 0.0  
30:00 215.6 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 21:58  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:01:22  
CYCLE COMPLETE

06/09/2023 10:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_