
Midmark M11 - v1.0.5
Total Cycles: 578

Name: RITTER2
Sterilizer ID: V2313228
11/25/2021 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.3 F 0.0
2:00 86.8 F 0.0
4:00 99.6 F 0.0
6:00 128.4 F 0.1
8:00 169.0 F 0.4
10:00 191.5 F 2.9
12:00 210.0 F 6.3
14:00 218.7 F 9.8
16:00 227.8 F 10.9
18:00 233.0 F 10.0
20:00 236.3 F 9.5
22:00 244.4 F 13.2
24:00 243.2 F 12.1
26:00 255.0 F 18.3
28:00 265.7 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.6 F 29.6
2:00 271.7 F 28.8
3:00 272.3 F 29.6
4:00 272.4 F 29.3

Min 271.3 F 28.7
Max 272.8 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.7 F 0.0
5:00 230.7 F 0.0
10:00 225.7 F 0.0
15:00 221.4 F 0.0
20:00 228.9 F 0.0
25:00 228.6 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 29:50
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:18
CYCLE COMPLETE

11/25/2021 04:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____