

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4752

Name: RITTER 1  
Sterilizer ID: V2178198  
07/22/2024 08:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.6 F 0.0  
2:00 69.0 F 0.0  
4:00 81.8 F 0.3  
6:00 111.1 F 2.4  
8:00 142.1 F 4.8  
10:00 167.1 F 7.6  
12:00 187.8 F 10.6  
14:00 199.6 F 4.1  
16:00 211.7 F 6.0  
18:00 222.1 F 9.5  
20:00 232.5 F 13.5  
22:00 233.9 F 9.2  
24:00 239.7 F 10.7  
26:00 243.6 F 12.2  
28:00 255.3 F 18.2  
30:00 265.9 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.2 F 28.6  
2:00 272.3 F 29.1  
3:00 271.4 F 28.4  
4:00 272.2 F 28.8

Min 271.4 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.4 F 0.0  
5:00 196.2 F 0.0  
10:00 196.5 F 0.0  
15:00 203.5 F 0.0  
20:00 213.9 F 0.0  
25:00 214.0 F 0.0  
30:00 212.9 F 0.0  
Drying Complete

FILLING: 0:45  
HEATING: 31:41  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:09:52  
CYCLE COMPLETE

07/22/2024 09:48 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_