
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001115
8:51 09/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 4:13
STERILIZING 5:38
136.0 C 220kPa 5:38
136.3 C 222kPa 6:08
136.3 C 222kPa 6:38
136.3 C 224kPa 7:08
136.4 C 224kPa 7:38
136.4 C 223kPa 8:08
136.3 C 223kPa 8:38
136.4 C 224kPa 9:08
136.4 C 224kPa 9:08
VENTING 9:08
AIR DRYING 10:03
CYCLE COMPLETE 25:03

Digital Signature #
EBA4BBE56C85576C
