
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002162
8:12 25/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:57
STERILIZING 3:00
136.4 C 222kPa 3:00
136.2 C 222kPa 3:30
136.2 C 223kPa 4:00
136.4 C 225kPa 4:30
136.2 C 223kPa 5:00
136.3 C 223kPa 5:30
136.4 C 226kPa 6:00
136.4 C 221kPa 6:30
136.4 C 221kPa 6:30
VENTING 6:30
AIR DRYING 7:21
CYCLE COMPLETE 27:58

Digital Signature #
90DAEDA64F327A63
