
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000513
12:07 21/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:08
STERILIZING 4:24
136.0 C 229kPa 4:24
136.3 C 233kPa 4:54
136.4 C 232kPa 5:24
136.4 C 233kPa 5:54
136.4 C 232kPa 6:24
136.4 C 233kPa 6:54
136.4 C 233kPa 7:24
136.4 C 232kPa 7:54
136.4 C 232kPa 7:54
VENTING 7:54
AIR DRYING 8:49
CYCLE COMPLETE 9:04

Digital Signature #
60EF3DF8240C4563
