

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001981  
14:28 26/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:43  
STERILIZING 3:36  
136.0 C 221kPa 3:36  
136.4 C 224kPa 4:06  
136.2 C 223kPa 4:36  
136.4 C 225kPa 5:06  
136.3 C 225kPa 5:36  
136.2 C 224kPa 6:06  
136.4 C 226kPa 6:36  
136.4 C 225kPa 7:06  
136.4 C 225kPa 7:06  
VENTING 7:06  
AIR DRYING 7:59  
CYCLE COMPLETE 12:20  
-----

Digital Signature #  
4C84A3C8E58F4993  
-----