
Midmark M11 - v1.0.0
Total Cycles: 5247

Name: RITTER 1
Sterilizer ID: V2178198
01/15/2025 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.8 F 0.0
2:00 98.6 F 0.0
4:00 110.3 F 0.0
6:00 144.0 F 0.2
8:00 176.3 F 2.6
10:00 201.6 F 6.2
12:00 210.5 F 10.0
14:00 224.8 F 9.3
16:00 225.8 F 7.8
18:00 236.3 F 12.3
20:00 235.5 F 9.1
22:00 241.6 F 10.8
24:00 245.4 F 12.8
26:00 256.6 F 18.7
28:00 266.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.3 F 28.8
2:00 271.7 F 28.2
3:00 271.6 F 28.4
4:00 272.2 F 28.8

Min 271.2 F 28.0
Max 272.3 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 204.9 F 0.0
10:00 214.9 F 0.0
15:00 227.1 F 0.0
20:00 231.4 F 0.0
25:00 228.6 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:38
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:08:14
CYCLE COMPLETE

01/15/2025 06:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____