
Midmark M11 - v1.0.0
Total Cycles: 4239

Name: RITTER 1
Sterilizer ID: V2178198
01/29/2024 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 90.1 F 0.2
4:00 117.2 F 1.0
6:00 113.2 F 3.2
8:00 143.2 F 5.6
10:00 165.8 F 8.6
12:00 182.2 F 12.2
14:00 202.9 F 2.9
16:00 211.9 F 6.0
18:00 223.9 F 9.5
20:00 234.2 F 13.4
22:00 234.3 F 9.0
24:00 241.5 F 11.6
26:00 243.0 F 11.9
28:00 255.1 F 17.9
30:00 265.7 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 28.7
2:00 271.4 F 28.3
3:00 272.2 F 28.6
4:00 271.4 F 28.4

Min 271.1 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 224.5 F 0.0
10:00 228.5 F 0.0
15:00 233.0 F 0.0
20:00 222.6 F 0.0
25:00 231.3 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 31:52
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 01:10:11
CYCLE COMPLETE

01/29/2024 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____