
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008514
9:25 04/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:49
STERILIZING 3:46
136.4 C 222kPa 3:46
136.4 C 221kPa 4:16
136.4 C 221kPa 4:46
136.1 C 222kPa 5:16
136.4 C 224kPa 5:46
136.4 C 225kPa 6:16
136.4 C 223kPa 6:46
136.4 C 223kPa 7:16
136.4 C 223kPa 7:16
VENTING 7:16
AIR DRYING 8:07
CYCLE COMPLETE 31:35

Digital Signature #
C64EB34F9BB1F29A
