

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000987  
13:21 10/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:05  
STERILIZING 3:06  
136.0 C 227kPa 3:06  
136.4 C 233kPa 3:36  
136.4 C 232kPa 4:06  
136.4 C 233kPa 4:36  
136.4 C 233kPa 5:06  
136.4 C 233kPa 5:36  
136.4 C 233kPa 6:06  
136.2 C 232kPa 6:36  
136.2 C 232kPa 6:36  
VENTING 6:36  
AIR DRYING 7:28  
CYCLE COMPLETE 7:34  
-----

Digital Signature #  
ED52C52B67E16666  
-----