
Midmark M11 - v1.0.0
Total Cycles: 5306

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2025 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.3 F 0.0
2:00 176.6 F 3.8
4:00 189.9 F 8.1
6:00 227.6 F 12.8
8:00 224.3 F 6.4
10:00 234.4 F 11.3
12:00 232.2 F 7.7
14:00 243.8 F 12.2
16:00 244.4 F 12.5
18:00 257.0 F 19.0
20:00 268.1 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.2 F 28.9
2:00 271.3 F 28.3
3:00 272.4 F 29.1
4:00 271.7 F 28.4

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 223.0 F 0.0
10:00 226.3 F 0.0
15:00 203.5 F 0.0
20:00 210.3 F 0.0
25:00 212.4 F 0.0
30:00 214.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:00:15
CYCLE COMPLETE

01/31/2025 03:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____