
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004003
13:56 13/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:49
PRESSURIZING 1:56
STERILIZING 2:47
136.0 C 220kPa 2:47
136.4 C 225kPa 3:17
136.4 C 224kPa 3:47
136.1 C 223kPa 4:17
136.3 C 223kPa 4:47
136.0 C 220kPa 5:17
136.1 C 222kPa 5:47
136.2 C 221kPa 6:17
136.2 C 221kPa 6:17
VENTING 6:17
AIR DRYING 8:08
CYCLE COMPLETE 12:00

Digital Signature #
A050B3F6DC5522C4
