

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2551

Name: RITTER 1  
Sterilizer ID: V2178198  
11/14/2022 12:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
2:00 179.6 F 3.9  
4:00 192.8 F 8.4  
6:00 210.9 F 13.5  
8:00 223.6 F 7.6  
10:00 237.3 F 12.9  
12:00 237.3 F 10.0  
14:00 238.0 F 8.5  
16:00 249.3 F 14.8  
18:00 261.3 F 21.7  
20:00 272.5 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.7 F 29.1  
2:00 271.9 F 28.5  
3:00 271.8 F 28.7  
4:00 272.7 F 29.1

Min 271.4 F 28.2  
Max 272.8 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 229.0 F 0.0  
10:00 233.1 F 0.0  
15:00 229.8 F 0.0  
20:00 224.5 F 0.0  
25:00 227.9 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:54  
CYCLE COMPLETE

11/14/2022 01:17 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_