

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4304

Name: RITTER2  
Sterilizer ID: V2313228  
01/28/2025 12:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 119.0 F 0.0  
2:00 118.9 F 0.7  
4:00 122.9 F 2.6  
6:00 142.2 F 5.5  
8:00 164.5 F 9.2  
10:00 186.6 F 13.4  
12:00 212.2 F 6.1  
14:00 225.8 F 10.3  
16:00 231.7 F 8.4  
18:00 237.9 F 11.3  
20:00 235.7 F 9.2  
22:00 247.7 F 14.9  
24:00 258.2 F 20.7  
26:00 267.8 F 26.9  
28:00 272.5 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 30.0  
1:00 271.9 F 29.9  
2:00 272.5 F 30.1  
3:00 272.3 F 30.4  
4:00 271.4 F 30.1

Min 271.3 F 29.8  
Max 272.7 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.3 F 0.0  
5:00 177.7 F 0.0  
10:00 176.4 F 0.0  
15:00 174.6 F 0.0  
20:00 170.4 F 0.0  
25:00 180.3 F 0.0  
30:00 179.0 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 28:10  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:06:55  
CYCLE COMPLETE

01/28/2025 01:47 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_