
Midmark M11 - v1.0.5
Total Cycles: 407

Name: RITTER2
Sterilizer ID: V2313228
09/29/2021 11:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.6 F 0.0
2:00 148.9 F 2.4
4:00 157.2 F 5.7
6:00 180.7 F 10.2
8:00 209.1 F 5.8
10:00 220.9 F 8.8
12:00 233.5 F 12.6
14:00 236.8 F 11.0
16:00 236.2 F 9.6
18:00 250.2 F 16.0
20:00 261.2 F 22.4
22:00 271.6 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.5 F 29.8
2:00 271.3 F 29.2
3:00 272.5 F 29.9
4:00 271.6 F 29.2

Min 271.2 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 214.9 F 0.0
10:00 212.6 F 0.0
15:00 206.9 F 0.0
20:00 209.2 F 0.0
25:00 210.7 F 0.0
30:00 215.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:01:07
CYCLE COMPLETE

09/29/2021 12:14 PM
ACCEPTED: X
REJECTED:
Operator: U01