
Midmark M11 - v1.0.5
Total Cycles: 2972

Name: RITTER2
Sterilizer ID: V2313228
10/04/2023 06:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.8 F 0.0
2:00 168.5 F 2.8
4:00 191.9 F 6.9
6:00 208.6 F 11.5
8:00 222.4 F 5.7
10:00 233.8 F 10.4
12:00 234.6 F 8.4
14:00 245.2 F 12.8
16:00 245.1 F 13.1
18:00 257.2 F 19.6
20:00 268.2 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.3 F 29.1
2:00 272.1 F 29.4
3:00 272.1 F 28.9
4:00 272.1 F 29.5

Min 271.3 F 28.6
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.3 F 0.0
5:00 226.0 F 0.0
10:00 224.4 F 0.0
15:00 199.5 F 0.0
20:00 184.7 F 0.0
25:00 178.2 F 0.0
30:00 174.3 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

10/04/2023 07:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____