
Midmark M11 - v1.0.0
Total Cycles: 4025

Name: RITTER 1
Sterilizer ID: V2178198
11/15/2023 04:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.6 F 0.0
2:00 141.4 F 2.1
4:00 155.8 F 5.3
6:00 181.5 F 9.1
8:00 212.2 F 9.1
10:00 213.7 F 6.6
12:00 226.0 F 11.1
14:00 228.7 F 6.9
16:00 241.4 F 12.4
18:00 239.7 F 10.5
20:00 252.3 F 16.6
22:00 263.2 F 22.9
24:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.4
1:00 271.5 F 28.7
2:00 272.6 F 29.3
3:00 272.2 F 28.9
4:00 271.5 F 28.4

Min 271.1 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 231.7 F 0.0
10:00 228.0 F 0.0
15:00 228.0 F 0.0
20:00 226.8 F 0.0
25:00 227.0 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 24:15
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

11/15/2023 05:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____