

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000015  
11:50 08/01/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:20  
STERILIZING 3:21  
136.4 C 225kPa 3:21  
136.0 C 223kPa 3:51  
136.4 C 225kPa 4:21  
136.3 C 221kPa 4:51  
136.4 C 224kPa 5:21  
136.4 C 220kPa 5:51  
136.2 C 222kPa 6:21  
135.9 C 219kPa 6:51  
135.9 C 219kPa 6:51  
VENTING 6:51  
AIR DRYING 7:40  
CYCLE COMPLETE 9:38

-----  
Digital Signature #  
12C25C7356426A7B  
-----