
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 000020
16:26 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 4:05
STERILIZING 5:54
136.4 C 227kPa 5:54
136.4 C 227kPa 6:24
136.4 C 227kPa 6:54
136.4 C 227kPa 7:24
136.5 C 228kPa 7:54
136.5 C 226kPa 8:24
136.3 C 226kPa 8:54
136.4 C 227kPa 9:24
VENTING 9:24
AIR DRYING 10:48
CYCLE COMPLETE 12:48

Digital Signature #
75B0FE363353186D
