

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000124  
15:52 04/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:15  
PRESSURIZING 3:19  
STERILIZING 4:31  
136.0 C 220kPa 4:31  
136.1 C 223kPa 5:01  
136.1 C 223kPa 5:31  
136.4 C 225kPa 6:01  
136.3 C 224kPa 6:31  
136.4 C 225kPa 7:01  
136.3 C 224kPa 7:31  
136.4 C 222kPa 8:01  
136.4 C 222kPa 8:01  
VENTING 8:02  
AIR DRYING 8:57  
CYCLE COMPLETE 21:03  
-----

Digital Signature #  
69666B24D3432EBD  
-----