
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000482
15:38 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:31
PRESSURIZING 2:39
STERILIZING 3:38
136.4 C 220kPa 3:38
136.4 C 224kPa 4:08
136.1 C 221kPa 4:38
136.1 C 220kPa 5:08
136.4 C 225kPa 5:38
136.4 C 225kPa 6:08
136.4 C 224kPa 6:38
136.5 C 224kPa 7:08
136.5 C 224kPa 7:08
VENTING 7:08
AIR DRYING 8:01
CYCLE COMPLETE 8:15

Digital Signature #
F6758DD89B58C3C0
