
Midmark M11 - v1.0.5
Total Cycles: 3

Name: RITTER2
Sterilizer ID: V2313228
04/21/2021 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.6 F 0.0
2:00 92.8 F 0.8
4:00 110.7 F 2.9
6:00 143.5 F 6.2
8:00 171.0 F 10.0
10:00 201.1 F 4.7
12:00 209.7 F 6.9
14:00 222.5 F 11.1
16:00 228.0 F 7.2
18:00 240.7 F 13.0
20:00 241.3 F 11.8
22:00 253.6 F 18.0
24:00 264.4 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 271.7 F 29.3
2:00 271.8 F 29.4
3:00 272.1 F 29.6
4:00 272.3 F 29.9

Min 271.6 F 29.3
Max 272.5 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 233.4 F 0.0
10:00 221.2 F 0.0
15:00 228.9 F 0.0
20:00 221.8 F 0.0
25:00 219.3 F 0.0
30:00 172.6 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:55
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:04:01
CYCLE COMPLETE

04/21/2021 11:55 AM
ACCEPTED: X
REJECTED:
Operator: U01