
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000730
15:59 29/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:08
STERILIZING 4:14
136.5 C 226kPa 4:14
136.4 C 226kPa 4:44
136.3 C 225kPa 5:14
136.2 C 221kPa 5:44
136.3 C 224kPa 6:14
136.3 C 221kPa 6:44
136.2 C 223kPa 7:14
136.2 C 222kPa 7:44
136.2 C 222kPa 7:44
VENTING 7:44
AIR DRYING 8:40
CYCLE COMPLETE 8:45

Digital Signature #
FC01DCE5A9139F6C
