
Midmark M11 - v1.0.0
Total Cycles: 819

Name: RITTER 1
Sterilizer ID: V2178198
05/19/2021 07:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 73.0 F 0.0
4:00 91.7 F 0.0
6:00 131.0 F 0.1
8:00 152.4 F 0.2
10:00 182.4 F 0.3
12:00 203.8 F 1.9
14:00 214.5 F 5.0
16:00 223.0 F 7.6
18:00 232.4 F 10.9
20:00 232.2 F 7.7
22:00 239.3 F 10.9
24:00 235.1 F 7.9
26:00 244.1 F 11.5
28:00 246.6 F 13.1
30:00 259.0 F 19.8
32:00 269.7 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.4 F 28.3
2:00 272.0 F 28.3
3:00 272.2 F 28.2
4:00 272.1 F 28.5

Min 271.6 F 27.8
Max 272.7 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.7 F 0.0
5:00 221.4 F 0.0
10:00 209.8 F 0.0
15:00 196.6 F 0.0
20:00 187.5 F 0.0
25:00 182.0 F 0.0
Dry Cycle Aborted

FILLING: 0:53
HEATING: 32:58
STERILIZING: 4:00
VENTING: 3:08
DRYING: 27:49
TOTAL CYCLE: 01:08:48

05/19/2021 09:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____