
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000836
15:55 31/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 4:40
STERILIZING 6:08
136.0 C 220kPa 6:08
136.4 C 223kPa 6:38
136.0 C 221kPa 7:08
136.4 C 223kPa 7:38
135.7 C 218kPa 8:08
136.4 C 223kPa 8:38
136.4 C 221kPa 9:08
136.3 C 223kPa 9:38
136.3 C 223kPa 9:38
VENTING 9:38
AIR DRYING 10:36
CYCLE COMPLETE 25:13

Digital Signature #
20C5E094270F47C7
