
Midmark M11 - v1.0.5
Total Cycles: 4240

Name: RITTER2
Sterilizer ID: V2313228
01/02/2025 05:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.9 F 0.0
2:00 99.5 F 0.0
4:00 123.0 F 0.2
6:00 154.8 F 3.4
8:00 180.9 F 7.0
10:00 197.6 F 10.9
12:00 212.1 F 4.8
14:00 225.0 F 9.1
16:00 234.4 F 11.3
18:00 238.5 F 11.1
20:00 236.5 F 9.3
22:00 249.9 F 15.7
24:00 261.2 F 22.2
26:00 271.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.7 F 29.5
2:00 272.0 F 29.4
3:00 272.4 F 29.3
4:00 272.2 F 29.6

Min 271.7 F 28.8
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.6 F 0.0
5:00 194.5 F 0.0
10:00 208.7 F 0.0
15:00 219.8 F 0.0
20:00 223.3 F 0.0
25:00 225.0 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 26:35
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:05:01
CYCLE COMPLETE

01/02/2025 06:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____