
Midmark M11 - v1.0.5
Total Cycles: 923

Name: RITTER2
Sterilizer ID: V2313228
03/29/2022 02:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.0 F 0.0
2:00 186.3 F 0.5
4:00 213.2 F 4.5
6:00 235.1 F 10.1
8:00 226.6 F 7.1
10:00 244.4 F 12.6
12:00 241.7 F 11.2
14:00 239.9 F 10.4
16:00 254.1 F 17.4
18:00 266.0 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 272.7 F 29.1
2:00 271.6 F 28.5
3:00 271.3 F 28.8
4:00 272.7 F 29.2

Min 271.0 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 232.4 F 0.0
10:00 231.7 F 0.0
15:00 229.4 F 0.0
20:00 230.7 F 0.0
25:00 229.0 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 19:38
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:60
CYCLE COMPLETE

03/29/2022 03:23 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____