

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3716

Name: RITTER2  
Sterilizer ID: V2313228  
06/19/2024 10:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 127.2 F 0.0  
2:00 130.3 F 1.3  
4:00 137.6 F 3.9  
6:00 162.4 F 7.4  
8:00 186.6 F 11.6  
10:00 209.4 F 4.3  
12:00 222.7 F 8.7  
14:00 234.3 F 13.5  
16:00 235.5 F 10.5  
18:00 235.1 F 8.4  
20:00 247.5 F 14.6  
22:00 258.7 F 20.8  
24:00 268.8 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 273.0 F 30.1  
2:00 271.7 F 29.3  
3:00 272.9 F 30.0  
4:00 271.6 F 29.4

Min 271.0 F 28.8  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.5 F 0.0  
5:00 201.3 F 0.0  
10:00 204.8 F 0.0  
15:00 192.5 F 0.0  
20:00 192.9 F 0.0  
25:00 194.8 F 0.0  
30:00 209.4 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 25:11  
STERILIZING: 4:00  
VENTING: 3:28  
DRYING: 30:00  
TOTAL CYCLE: 01:03:33  
CYCLE COMPLETE

06/19/2024 11:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_