
Midmark M11 - v1.0.0
Total Cycles: 2597

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.5 F 0.0
2:00 166.5 F 3.5
4:00 186.1 F 8.0
6:00 208.5 F 13.2
8:00 223.5 F 7.7
10:00 238.0 F 13.4
12:00 239.2 F 11.0
14:00 238.0 F 9.7
16:00 253.0 F 17.0
18:00 265.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.6 F 29.1
2:00 271.8 F 28.3
3:00 272.8 F 29.3
4:00 272.3 F 28.7

Min 271.7 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.7 F 0.0
5:00 217.7 F 0.0
10:00 238.9 F 0.0
15:00 228.6 F 0.0
20:00 219.3 F 0.0
25:00 213.2 F 0.0
30:00 209.6 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:40
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

11/22/2022 03:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____