
Midmark M11 - v1.0.0
Total Cycles: 984

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2021 11:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.0 F 0.0
2:00 170.6 F 3.1
4:00 193.1 F 7.1
6:00 215.7 F 12.7
8:00 227.9 F 7.2
10:00 240.9 F 12.4
12:00 237.7 F 9.5
14:00 237.4 F 8.8
16:00 249.4 F 14.7
18:00 261.6 F 21.6
20:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.1
1:00 272.5 F 28.6
2:00 272.2 F 28.7
3:00 271.5 F 28.0
4:00 272.2 F 28.5

Min 271.4 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 197.5 F 0.0
10:00 203.3 F 0.0
15:00 206.9 F 0.0
20:00 205.5 F 0.0
25:00 210.4 F 0.0
30:00 239.9 F 0.0
Drying Complete

FILLING: 2:02
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

07/08/2021 12:16 PM
ACCEPTED: X
REJECTED:
Operator: U02