
Midmark M11 - v1.0.5
Total Cycles: 917

Name: RITTER2
Sterilizer ID: V2313228
03/28/2022 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.1 F 0.0
2:00 159.5 F 0.1
4:00 189.2 F 2.1
6:00 205.9 F 7.0
8:00 221.7 F 12.3
10:00 229.2 F 8.9
12:00 238.0 F 10.3
14:00 244.6 F 13.5
16:00 245.2 F 13.3
18:00 258.4 F 20.4
20:00 270.1 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.6
2:00 271.4 F 28.9
3:00 272.7 F 29.7
4:00 272.1 F 29.1

Min 271.2 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 227.5 F 0.0
10:00 216.9 F 0.0
15:00 220.8 F 0.0
20:00 222.2 F 0.0
25:00 220.7 F 0.0
30:00 222.5 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:17
CYCLE COMPLETE

03/28/2022 04:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____