
Midmark M11 - v1.0.5
Total Cycles: 4313

Name: RITTER2
Sterilizer ID: V2313228
01/29/2025 04:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 143.4 F 0.0
2:00 157.4 F 1.7
4:00 149.0 F 5.0
6:00 174.3 F 9.1
8:00 193.5 F 13.5
10:00 216.3 F 6.4
12:00 226.7 F 10.4
14:00 232.0 F 8.7
16:00 239.0 F 11.9
18:00 238.2 F 10.4
20:00 250.7 F 16.5
22:00 261.5 F 22.8
24:00 271.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.6
1:00 272.6 F 30.4
2:00 272.0 F 29.9
3:00 271.9 F 30.4
4:00 272.7 F 30.4

Min 270.9 F 29.5
Max 273.6 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 195.2 F 0.0
10:00 215.3 F 0.0
15:00 203.4 F 0.0
20:00 203.5 F 0.0
25:00 204.4 F 0.0
30:00 199.6 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:42
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:03:42
CYCLE COMPLETE

01/29/2025 05:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____