
Midmark M11 - v1.0.0
Total Cycles: 4878

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2024 10:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.5 F 0.0
2:00 155.7 F 0.1
4:00 160.8 F 0.3
6:00 183.9 F 0.7
8:00 204.1 F 2.7
10:00 213.1 F 6.3
12:00 223.7 F 10.1
14:00 233.9 F 9.3
16:00 236.3 F 10.3
18:00 236.6 F 8.8
20:00 246.2 F 13.6
22:00 244.7 F 12.5
24:00 256.2 F 18.5
26:00 266.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.1 F 28.6
2:00 272.6 F 28.9
3:00 272.4 F 28.7
4:00 272.1 F 28.4

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 216.1 F 0.0
10:00 221.3 F 0.0
15:00 222.1 F 0.0
20:00 223.4 F 0.0
25:00 225.5 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 27:56
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:06:37
CYCLE COMPLETE

09/09/2024 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____