
Midmark M11 - v1.0.5
Total Cycles: 1972

Name: RITTER2
Sterilizer ID: V2313228
01/25/2023 11:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.1 F 0.0
2:00 192.8 F 3.6
4:00 200.0 F 8.8
6:00 220.1 F 8.5
8:00 230.1 F 10.2
10:00 233.1 F 8.0
12:00 241.4 F 11.4
14:00 246.3 F 13.8
16:00 259.3 F 20.8
18:00 270.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.1 F 29.1
2:00 271.8 F 29.5
3:00 271.8 F 28.9
4:00 272.1 F 29.5

Min 271.4 F 28.7
Max 272.5 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 223.7 F 0.0
10:00 227.7 F 0.0
15:00 234.2 F 0.0
20:00 232.5 F 0.0
25:00 223.1 F 0.0
30:00 221.5 F 0.0

Drying Complete

FILLING: 1:45
HEATING: 18:47
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:44
CYCLE COMPLETE

01/25/2023 12:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____