
Midmark M11 - v1.0.0
Total Cycles: 3773

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2023 12:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.6 F 0.0
2:00 168.3 F 2.2
4:00 184.1 F 5.9
6:00 205.2 F 10.3
8:00 226.5 F 6.6
10:00 228.5 F 9.8
12:00 235.0 F 8.9
14:00 242.0 F 12.2
16:00 240.7 F 10.6
18:00 253.2 F 16.9
20:00 263.9 F 23.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.6 F 29.2
2:00 271.6 F 28.7
3:00 272.2 F 28.7
4:00 272.7 F 29.1

Min 271.0 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 219.8 F 0.0
10:00 216.6 F 0.0
15:00 215.4 F 0.0
20:00 201.4 F 0.0
25:00 202.6 F 0.0
30:00 189.9 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:55
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:00:53
CYCLE COMPLETE

09/14/2023 01:33 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____