

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2395

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2022 10:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
2:00 185.6 F 3.4  
4:00 191.2 F 7.2  
6:00 214.6 F 12.7  
8:00 226.3 F 7.1  
10:00 238.0 F 12.4  
12:00 236.4 F 9.6  
14:00 239.1 F 8.8  
16:00 248.9 F 14.7  
18:00 261.1 F 21.5  
20:00 272.0 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.6 F 29.0  
2:00 271.5 F 28.4  
3:00 272.7 F 29.1  
4:00 271.7 F 28.3

Min 271.5 F 28.2  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 214.0 F 0.0  
10:00 229.2 F 0.0  
15:00 236.6 F 0.0  
20:00 229.4 F 0.0  
25:00 224.0 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:16  
HEATING: 20:32  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

10/04/2022 11:42 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_