
Midmark M11 - v1.0.5
Total Cycles: 1334

Name: RITTER2
Sterilizer ID: V2313228
07/25/2022 07:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.7 F 0.0
2:00 74.6 F 0.0
4:00 109.3 F 0.0
6:00 138.7 F 0.0
8:00 164.3 F 1.0
10:00 181.8 F 4.0
12:00 196.6 F 7.2
14:00 209.4 F 10.5
16:00 221.1 F 8.4
18:00 227.7 F 7.8
20:00 237.6 F 11.9
22:00 233.6 F 8.6
24:00 241.8 F 12.1
26:00 242.9 F 12.4
28:00 254.0 F 18.2
30:00 264.1 F 24.4
32:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.4 F 29.9
2:00 271.4 F 29.1
3:00 272.4 F 30.0
4:00 271.9 F 29.4

Min 271.2 F 29.1
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 211.5 F 0.0
10:00 226.1 F 0.0
15:00 231.2 F 0.0
20:00 234.4 F 0.0
25:00 223.8 F 0.0
30:00 220.4 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 32:10
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:10:05
CYCLE COMPLETE

07/25/2022 08:29 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____