
Midmark M11 - v1.0.0
Total Cycles: 3614

Name: RITTER 1
Sterilizer ID: V2178198
08/03/2023 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.3 F 0.0
2:00 183.2 F 3.4
4:00 196.9 F 7.7
6:00 216.6 F 13.4
8:00 226.6 F 8.0
10:00 238.7 F 13.2
12:00 238.3 F 10.6
14:00 236.5 F 8.6
16:00 250.0 F 15.5
18:00 262.0 F 22.2
20:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.5
1:00 272.5 F 29.2
2:00 271.3 F 28.5
3:00 272.4 F 29.0
4:00 271.3 F 28.5

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.0 F 0.0
5:00 218.0 F 0.0
10:00 229.1 F 0.0
15:00 234.0 F 0.0
20:00 223.0 F 0.0
25:00 228.4 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:13
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

08/03/2023 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____