
Midmark M11 - v1.0.5
Total Cycles: 1727

Name: RITTER2
Sterilizer ID: V2313228
11/15/2022 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.4 F 0.0
2:00 194.0 F 3.4
4:00 201.2 F 9.2
6:00 219.3 F 7.3
8:00 233.3 F 11.2
10:00 234.9 F 9.3
12:00 239.0 F 9.3
14:00 250.4 F 16.0
16:00 263.4 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 272.8 F 29.6
2:00 271.8 F 29.1
3:00 272.8 F 29.6
4:00 271.9 F 29.0

Min 271.7 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 208.6 F 0.0
10:00 225.7 F 0.0
15:00 232.8 F 0.0
20:00 226.6 F 0.0
25:00 213.3 F 0.0
30:00 206.6 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 17:57
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

11/15/2022 02:15 PM
ACCEPTED: X
REJECTED:
Operator: U02