
Midmark M11 - v1.0.5
Total Cycles: 579

Name: RITTER2
Sterilizer ID: V2313228
11/29/2021 07:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 83.1 F 0.0
4:00 109.3 F 0.0
6:00 112.4 F 0.1
8:00 151.1 F 0.2
10:00 179.3 F 0.4
12:00 200.8 F 2.4
14:00 217.5 F 5.9
16:00 225.5 F 9.4
18:00 235.4 F 13.3
20:00 233.8 F 9.0
22:00 243.7 F 12.5
24:00 241.8 F 11.7
26:00 239.8 F 10.4
28:00 253.8 F 17.5
30:00 265.7 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.8 F 29.3
2:00 271.7 F 28.6
3:00 272.9 F 29.3
4:00 271.8 F 28.5

Min 271.6 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 215.5 F 0.0
10:00 204.8 F 0.0
15:00 209.9 F 0.0
20:00 217.9 F 0.0
25:00 218.7 F 0.0
30:00 219.9 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 31:42
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:09:47
CYCLE COMPLETE

11/29/2021 10:07 AM
ACCEPTED: X
REJECTED:
Operator: U02