

-----  
Midmark M11 - v1.0.5  
Total Cycles: 530

Name: RITTER2  
Sterilizer ID: V2313228  
11/09/2021 02:27 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
2:00 200.6 F 4.2  
4:00 203.5 F 9.8  
6:00 221.7 F 6.9  
8:00 235.0 F 12.3  
10:00 237.3 F 10.7  
12:00 238.1 F 10.0  
14:00 252.5 F 17.2  
16:00 264.9 F 24.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.2 F 29.3  
2:00 272.5 F 29.7  
3:00 271.8 F 29.6  
4:00 272.0 F 29.3  
  
Min 271.5 F 29.0  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 228.3 F 0.0  
10:00 239.9 F 0.0  
15:00 228.7 F 0.0  
20:00 227.3 F 0.0  
25:00 226.5 F 0.0  
30:00 224.2 F 0.0  
Drying Complete

FILLING: 2:57  
HEATING: 17:41  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:47  
CYCLE COMPLETE

11/09/2021 03:24 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_