
Midmark M11 - v1.0.5
Total Cycles: 791

Name: RITTER2
Sterilizer ID: V2313228
02/14/2022 07:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.5 F 0.0
2:00 77.9 F 0.0
4:00 106.5 F 0.0
6:00 111.8 F 0.4
8:00 150.8 F 3.3
10:00 174.1 F 6.5
12:00 193.4 F 10.3
14:00 216.4 F 7.8
16:00 219.4 F 6.7
18:00 230.1 F 10.6
20:00 234.2 F 8.3
22:00 241.4 F 12.5
24:00 240.3 F 11.1
26:00 252.4 F 17.2
28:00 263.0 F 23.5
30:00 272.6 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.0 F 29.2
2:00 272.0 F 29.5
3:00 272.0 F 29.8
4:00 271.5 F 29.7

Min 271.3 F 29.1
Max 272.2 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 201.7 F 0.0
10:00 197.5 F 0.0
15:00 195.8 F 0.0
20:00 195.4 F 0.0
25:00 198.5 F 0.0
30:00 202.6 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 30:14
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:08:22
CYCLE COMPLETE

02/14/2022 08:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____